

Hamilton Bonaduz AG, CH-7402 Bonaduz / Switzerland confirms that the following products

Product Families **VisiFerm mA ***** with SN 1000 and higher
VisiTrace mA *** with SN 1000 and higher

Note: *** represents the a-length of the sensor (120 to 600 mm)

meets the following EU directives (including all applicable amendments):



0035

EMC Directive	2014/30/EU
RED Directive	2014/53/EU
RoHS 2 Directive	2011/65/EU + 2015/863/EU
ATEX Directive	2014/34/EU
EC type Examination Certificate	TÜV 18 ATEX 8236X

Applied company quality management systems

EN ISO 9001: 2015
 IEC 80079-34: 2018

Certification Body

TÜV Rheinland Industrie Service GmbH
 Am Grauen Stein 29
 D-51105 Köln-Poll, Germany

Applied standards

EMC	EN 61326-1:2013, Emission Class B, Immunity Industrial environnement
RED	EN 300 328 V2.2.2:2019 EN 301 489-1 V2.2.0:2017 EN 301 489-17 V3.2.0:2017
ATEX	EN 60079-0:2018; IEC 60079-0: 2017 (ed.7.0) EN 60079-11:2012; IEC 60079-11:2011 (ed. 6.0)
RoHS	EN 63000:2019

Additional information

IECEX The products are covered by IECEX Certificate **IECEX TUR 19.0050X**.

China RoHS "Restricted Use of Hazardous Substances in Electronic and Electrical Products", the products are marked with , see hamiltoncompany.com for hazardous substances.

Wireless approvals

	USA	Canada	Japan	China	India	South-Korea
	FCC ID	IC ID	Certification No.	CMIIT ID	Registration No.	Registration No.
VisiFerm mA VisiTrace mA	2AQYJVISIMA	24225-VISIMA	203-JN1128	2021DJ3198 2021DJ3143	ETA-SD-20200704294 ETA-SD-20200704297	R-R-HBz-VmA

Authorize department for
 technical documents:

HAMILTON Process Analytics

Hamilton Bonaduz AG

Dr. Philipp Arquint
 Vice President Process Analytics

Bonaduz, **2024-04-08 170 PhA**