

EU Declaration of Conformity for Transmitter H110

We Hamilton Bonaduz AG, CH-7402 Bonaduz/Switzerland confirm that the following product

Product name **H110 Transmitter** Ref 10174805-xx

Note 1: xx represents the combination of two possible measurement channels for pH, Dissolved Oxygen or Conductivity measurements.

meets the following EU directives (including all applicable amendments):



Low Voltage Directive	2014/35/EU
EMC Directive	2014/30/EU
RoHS 2 Directive	2011/65/EU

Applied company quality management systems
EN ISO 9001:2015

Certification Body
TÜV Rheinland Industrie Service GmbH
Am Grauen Stein 29
D-51105 Köln-Poll, Germany

Applied standards:

Low Voltage Directive: EN 61010-1: 2010:2015

EMC Directive: EN 61326-1: 2013

RoHS 2 Directive: IEC 63000:2018

Additional Information:



Canada, USA
Certificate No. 80215882
Master Contract: 607077

CSA C22.2 No. 25-1966
CSA C22.2 No. 94.2:20
CSA C22.2 No. 213-M1987
CAN/CSA-C22.2 NO. 61010-1-12 AMD1
CSA C22.2 No. 60529:05
(Reaffirmed 2015)
ANSI/ISA 12.12.01: 2007
UL Standard 50E, 3rd Edition
UL 61010-1, 3rd Edition

Authorize department for
technical documents:

HAMILTON Process Analytics

Hamilton Bonaduz AG

Philipp Arquint
Vice President Process Analytics

Bonaduz, **2024 -10- 10 170 PhA**