

EU Declaration of Conformity for Cell Density Monitoring System

We Hamilton Bonaduz AG, CH-7402 Bonaduz/Switzerland confirm that the following products

Product name	Ref	Comment
Arc View 265 Controller	243800	
Arc View 465 Controller	243801	
Arc View 465 XL Controller	243802	
Incyte pre-amp	243720	
Dencytee pre-amp	243760	
4-20 mA Output Box	243850	
Incyte Sensor Family	various	
Dencytee Sensor Family	various	
Cell Density Integration Kit	243809	
Cell Density ComBox	243810	

meet the following EU directives (including all applicable amendments):



RoHS 2 Directive	2011/65/EU
EMC Directive	2014/30/EU
Low Voltage Directive	2014/35/EU

Applied company quality management system
EN ISO 9001

Certification Body
TÜV Rheinland Industrie Service GmbH
Am Grauen Stein 29
D-51105 Köln-Poll, Germany

Applied standards:

RoHS 2 Directive:	EN 50581:2012
EMC Directive:	EN 61326-1:2013 EN 61000-3-2:2006 + A1:2009 + A2:2009 EN 61000-3-3:2008 EN 61000-6-2 / EN 61000-6-4 (243809 only)
Low Voltage Directive:	EN 60950-1

Additional information:

All products listed in this document comply with CFR47 FCC Part 15.

Authorize department for
technical documents:

HAMILTON Process Analytics

Hamilton Bonaduz AG

Andreas Wieland
CEO

Bonaduz, 12.7.2017